

# SARS-COV-2 MOLECULAR ASSAY EVALUATION: RESULTS

INFORMATION FROM [WWW.FINDDX.ORG/COVID-19/SARSCOV2-EVAL-MOLECULAR/MOLECULAR-EVAL-RESULTS/](http://WWW.FINDDX.ORG/COVID-19/SARSCOV2-EVAL-MOLECULAR/MOLECULAR-EVAL-RESULTS/)  
LAST UPDATED: 15 APRIL 2020

FIND conducted independent evaluations at the [Hôpitaux Universitaires de Genève](https://www.hug.ch/), to verify the limit of detection (LOD) and the clinical performance (as reported by the manufacturers) of the following molecular test kits. The LOD analysis was performed using cultured viral stocks from a clinical isolate from Switzerland, and quantified using an E gene standard. The clinical performance analysis was conducted on extracted samples from individuals suspected to have COVID-19 that were tested using an in-house PCR protocol that was optimized based on the Tib Molbiol assay.

Data for the first tests evaluated are summarized below. More results will be added as they become available. Tests were selected for evaluation according to the [scoring criteria](#), but the order in which the evaluations were conducted does not reflect any endorsement or prioritization.

INFORMATION FROM [WWW.FINDDX.ORG/COVID-19/SARSCOV2-EVAL-MOLECULAR/MOLECULAR-EVAL-RESULTS/](http://WWW.FINDDX.ORG/COVID-19/SARSCOV2-EVAL-MOLECULAR/MOLECULAR-EVAL-RESULTS/)  
LAST UPDATED: 15 APRIL 2020

Company	Gene target	Verified LOD (copies / reaction)	Avg Ct (lowest dilution 10/10)	Clinical sensitivity (50 positives)	Clinical specificity* (100 negatives)	Product No.	Product name	Lot No.	PCR platform	Supplier recommended Ct cut-off
KH Medical Co. Ltd	S	1–10	37.94	100% (95%CI: 93, 100)	100% (95%CI: 96, 100)	RV008	RADI COVID-19 Detection Kit	V008.200202	BioRad CFX96 deep well	≤ 40
	RdRP	10–50	36.74	100% (95%CI: 93, 100)	100% (95%CI: 96, 100)					
SD Biosensor Inc.	E	1–10	37.43	100% (95%CI: 93, 100)	97% (95%CI: 92, 99)	M-NCOV-01	STANDARD M nCoV Real-Time Detection Kit	MNC00120005	Roche LightCycler 480	≤ 41
	ORF1	1–10	36.99	100% (95%CI: 93, 100)	99% (95%CI: 95, 100)					
BGI Health (HK) Co. Ltd	ORF1	1–10	32.43	100% (95%CI: 93, 100)	99% (95%CI: 95, 100)	MFG030010	Real-time Fluorescent RT-PCR kit for detection 2019-nCoV (CE-IVD)	6220200305	Roche LightCycler 480	≤38
DAAN Gene Co. Ltd	ORF1	1–10	38.76	100% (95%CI: 93, 100)	96% (95%CI: 90, 98)	DA0930-DA0932	Detection Kit for 2019 Novel Coronavirus (2019-nCoV) RNA (PCR-Fluorescence Probing)	2020007	Roche LightCycler 480	≤ 40
	N	1–10	36.97	100% (95%CI: 93, 100)	98% (95%CI: 93, 99)					
Tib Molbiol	E	1–10	33.34	–	–	53-0776-96 6754155001	ModularDx Kit SARS-CoV (COVID19) E-gene (Tib Molbiol) + LightCycler Multiplex RNA Virus Master (Roche)	48202019	Roche LightCycler 480	Define the cut-off 2–4 cycles higher than observed Cp value for 10 copies

\* Further investigation needed to determine if apparent false positives are truly false positives or whether they are a due to a false negative reference standard result.

